

## Central Lancashire Online Knowledge (CLOK)

Title	Early outcomes of the novel Myval THV series compared to SAPIEN THV series and Evolut THV series in individuals with severe aortic stenosis
Type	Article
URL	<a href="https://clock.uclan.ac.uk/53873/">https://clock.uclan.ac.uk/53873/</a>
DOI	<a href="https://doi.org/10.4244/EIJ-D-24-00951">https://doi.org/10.4244/EIJ-D-24-00951</a>
Date	2025
Citation	van Royen, Niels, Amat-Santos, Ignacio J., Hudec, Martin, Bunc, Matjaz, Ijsselmuiden, Alexander, Laanmets, Peep, Unic, Daniel, Merkely, Béla, Hermanides, Renicus S. et al (2025) Early outcomes of the novel Myval THV series compared to SAPIEN THV series and Evolut THV series in individuals with severe aortic stenosis. <i>EuroIntervention</i> , 21 (2).
Creators	van Royen, Niels, Amat-Santos, Ignacio J., Hudec, Martin, Bunc, Matjaz, Ijsselmuiden, Alexander, Laanmets, Peep, Unic, Daniel, Merkely, Béla, Hermanides, Renicus S., Ninios, Vlasia, Protasiewicz, Marcin, Rensing, Benno J. W. M., Martin, Pedro L., Feres, Fausto, De Sousa, Manuel, Van Belle, Eric, Linke, Axel, Ielasi, Alfonso, Montofarno, Matteo, Webster, Mark, Toutouzas, Konstantinos, Teiger, Emmanuel, Bedogni, Francesco, Voskuil, Michiel, Pan, Manuel, Angerås, Oskar, Kim, Won-Keun, Rothe, Jürgen, Kristic, Ivica, Peral, Vicente, Van den Branden, Ben J. L., Westermann, Dirk, Bellini, Barbara, Garcia-Gomez, Mario, Tobe, Akihiro, Tsai, Tsung-Ying, Garg, Scot, Thakkar, Ashokkumar, Chandra, Udit, Morice, Marie-Claude, Soliman, Osama, Onuma, Yoshinobu, Serruys, Patrick W. and Baumbach, Andreas

It is advisable to refer to the publisher's version if you intend to cite from the work.  
<https://doi.org/10.4244/EIJ-D-24-00951>

For information about Research at UCLan please go to <http://www.uclan.ac.uk/research/>

All outputs in CLOK are protected by Intellectual Property Rights law, including Copyright law. Copyright, IPR and Moral Rights for the works on this site are retained by the individual authors and/or other copyright owners. Terms and conditions for use of this material are defined in the <http://clock.uclan.ac.uk/policies/>

# Early outcomes of the novel Myval THV series compared to SAPIEN THV series and Evolut THV series in individuals with severe aortic stenosis

Niels van Royen<sup>1</sup>, MD, PhD; Ignacio J. Amat-Santos<sup>2,3</sup>, MD, PhD; Martin Hudec<sup>4</sup>, MD, PhD; Matjaz Bunc<sup>5</sup>, MD, PhD; Alexander Ijsselmuiden<sup>6-8</sup>, MD, PhD; Peep Laanmets<sup>9</sup>, MD; Daniel Unic<sup>10</sup>, MD, PhD; Bela Merkely<sup>11</sup>, MD; Renicus S. Hermanides<sup>12</sup>, MD, PhD; Vlasios Ninios<sup>13</sup>, MD; Marcin Protasiewicz<sup>14</sup>, MD, PhD; Benno J.W.M. Rensing<sup>15</sup>, MD, PhD; Pedro L. Martin<sup>16</sup>, MD, PhD; Fausto Feres<sup>17</sup>, MD; Manuel De Sousa Almeida<sup>18</sup>, MD, PhD; Eric van Belle<sup>19</sup>, MD, PhD; Axel Linke<sup>20</sup>, MD; Alfonso Ielasi<sup>21</sup>, MD; Matteo Montorfano<sup>22,23</sup>, MD; Mark Webster<sup>24</sup>, MBChB; Konstantinos Toutouzas<sup>25</sup>, MD, PhD; Emmanuel Teiger<sup>26</sup>, MD, PhD; Francesco Bedogni<sup>27</sup>, MD; Michiel Voskuil<sup>28</sup>, MD, PhD; Manuel Pan<sup>29</sup>, MD, PhD; Oskar Angerås<sup>30,31</sup>, MD, PhD; Won-Keun Kim<sup>32,33</sup>, MD; Jürgen Rothe<sup>34,35</sup>, MD; Ivica Kristić<sup>36</sup>, MD, PhD; Vicente Peral<sup>37</sup>, MD, PhD; Ben J.L. Van den Branden<sup>6</sup>, MD; Dirk Westermann<sup>34</sup>, MD; Barbara Bellini<sup>23</sup>, MD; Mario Garcia-Gomez<sup>2,3</sup>, MD; Akihiro Tobe<sup>38</sup>, MD; Tsung-Ying Tsai<sup>38,39</sup>, MD; Scot Garg<sup>40,41</sup>, MD, PhD; Ashokkumar Thakkar<sup>42</sup>, PhD; Udit Chandra<sup>42</sup>, PhD; Marie-Claude Morice<sup>43,44</sup>, MD, PhD; Osama Soliman<sup>38</sup>, MD, PhD; Yoshinobu Onuma<sup>38,45</sup>, MD, PhD; Patrick W. Serruys<sup>38\*</sup>, MD, PhD; Andreas Baumbach<sup>46,47</sup>, MD, PhD

N. van Royen and I.J. Amat-Santos contributed equally.

\*Corresponding author: Department of Cardiology, University of Galway, University Road, Galway, H91 TK33, Ireland.

E-mail: [patrick.w.j.c.serruys@gmail.com](mailto:patrick.w.j.c.serruys@gmail.com)

This paper also includes supplementary data published online at: <https://eurointervention.pcronline.com/doi/10.4244/EIJ-D-24-00951>

## ABSTRACT

**BACKGROUND:** There are limited head-to-head randomised trials comparing the performance of different transcatheter heart valves (THVs).

**AIMS:** We aimed to evaluate the non-inferiority of the balloon-expandable Myval THV series compared to the balloon-expandable SAPIEN THV series or the self-expanding Evolut THV series.

**METHODS:** The LANDMARK trial randomised 768 patients in a 1:1 ratio, (Myval THV series [n=384] vs contemporary series with 50% SAPIEN THV series [n=192] and 50% Evolut THV series [n=192]). The non-inferiority of Myval over the SAPIEN or Evolut THV series in terms of the 30-day primary composite safety and effectiveness endpoint as per the third Valve Academic Research Consortium (VARC-3) was tested in an intention-to-treat population with a predefined statistical power of 80% (1-sided alpha of 5%) for a non-inferiority margin of 10.44%.

**RESULTS:** The Myval THV series achieved non-inferiority for the primary composite endpoint over the SAPIEN THV series (24.7% vs 24.1%, risk difference [95% confidence interval {CI}]: 0.6% [not applicable {NA}] to 8.0]; p=0.0033) and the Evolut THV series (24.7% vs 30.0%, risk difference [95% CI]: -5.3% [NA to 2.5]; p<0.0001). The incidences of pacemaker implantation were comparable (Myval THV series: 15.0%, SAPIEN THV series: 17.3%, Evolut THV series: 16.8%). At 30 days, the mean pressure gradient and effective orifice area were significantly better with the Myval THV series compared to the SAPIEN THV series (p<0.0001) and better with the Evolut THV series than with the Myval THV series (p<0.0001). At 30 days, the proportion of moderate to severe prosthetic valve regurgitation was numerically higher with the Evolut THV series compared to the Myval THV series (7.4% vs 3.4%; p=0.06), while not significantly different between the Myval THV series and the SAPIEN THV series (3.4% vs 1.6%; p=0.32).

**CONCLUSIONS:** The Myval THV series is non-inferior to the SAPIEN THV series and the Evolut THV series in terms of the primary composite endpoint at 30 days. Clinical trial registration: ClinicalTrials.gov: NCT04275726; EudraCT number 2020-000,137-40.

**KEYWORDS:** aortic stenosis; balloon-expandable valve; non-inferiority; randomised trial; self-expanding valve; transcatheter heart valve

The LANDMARK trial was the first prospective, randomised controlled trial to show the non-inferiority of a novel platform, the balloon-expandable (BE) Myval (Meril Life Sciences) transcatheter heart valve (THV) series, over contemporary THV series (combined BE SAPIEN [Edwards Lifesciences] and self-expanding [SE] Evolut [Medtronic] THV series)<sup>1</sup>. In the main analysis, the early clinical outcomes of the Myval THV series were only compared with the combined control group (the Evolut and SAPIEN THV series) with no individual head-to-head comparisons reported<sup>1</sup>. In the present subanalysis, we report the prespecified, statistically powered comparison between the three individual arms, to provide more granularity to the outcomes. The analytical plan and statistical design enabled us to separately test for the non-inferiority of the Myval THV series against the SAPIEN or Evolut THV series in a separate fashion. Notably, this is also the first randomised comparison of two BE valve (BEV) technologies.

## Methods

### STUDY DESIGN AND PARTICIPANTS

This is a powered, predefined substudy of the LANDMARK trial, with an aim to individually assess the non-inferiority of the Myval THV series over the SAPIEN THV series and Evolut THV series for the primary composite safety and effectiveness endpoint at 30-day follow-up.

The LANDMARK trial (ClinicalTrials.gov: NCT04275726) was a prospective, non-inferiority, randomised, open-label trial conducted at 31 centres in 16 countries. It was designed to evaluate the primary safety and effectiveness endpoint reported at 30 days according to the third Valve Academic Research Consortium (VARC-3)<sup>2</sup>, as well as clinical and haemodynamic outcomes. The trial design, protocol amendment on eligibility criteria, and the main study results have been published previously<sup>1,3,4</sup>. The study was approved by the ethics committees of the respective study sites. Prior to screening, all study participants provided written informed consent. The study was carried out in accordance with the International Council for Harmonization of Technical Requirements for Pharmaceuticals for Human Use guidelines.

Adults (≥18 years old) with severe symptomatic native aortic stenosis (AS) were deemed suitable for enrolment if judged by the local Heart Team to be eligible for transcatheter aortic valve implantation (TAVI) utilising all three study devices. A prescreening committee assessed the appropriateness of TAVI with the study THVs using the preprocedural multislice computed tomography (MSCT) provided by an independent core lab (TAVI Core Lab, India). Each site's Heart Team made the ultimate decision about each subject's enrolment into the study.

## Impact on daily practice

The LANDMARK trial demonstrates that the Myval transcatheter heart valve (THV) series is non-inferior to both the SAPIEN and Evolut THV series in terms of safety and effectiveness at 30 days. One-year outcomes of the LANDMARK trial as well as the Compare-TAVI trial (Myval vs SAPIEN) will be available soon and are eagerly awaited.

The Myval THV series included Myval and Myval Octacor THVs with sizes 20 mm, 21.5 mm, 23 mm, 24.5 mm, 26 mm, 27.5 mm, and 29 mm in diameter. The SAPIEN THV series consisted of the SAPIEN 3 and SAPIEN 3 Ultra THVs, which were commercially available at the sites, with device sizes 20 mm, 23 mm, 26 mm, and 29 mm in diameter. Lastly, the Evolut THV series included the Evolut R and Evolut PRO THVs or any subsequent advanced commercial version available at the sites with sizes 23 mm, 26 mm, 29 mm, and 34 mm in diameter.

### RANDOMISATION AND MASKING

In this trial, 768 severe symptomatic native AS patients were enrolled in a 1:1 ratio to the Myval THV series (n=384) and contemporary THV series (n=384), with subsequent stratification and an equal allocation (1:1) of patients in the contemporary arm between the SAPIEN (n=192) and Evolut (n=192) THV series. A covariate-adaptive randomisation process was used based on the simulation in accordance with the Frane method, considering the power and selection bias concurrently<sup>5</sup>.

### PROCEDURES

The procedural details have been discussed in detail in the earlier publications<sup>1,3,4</sup>. In brief, preprocedural assessment included physical examination, medical history, laboratory investigations, electrocardiography (ECG), echocardiography, and MSCT. The protocol recommended a transfemoral approach. The implantation technique, use of sedation, the need for predilation and post-dilation, and closure of the femoral access (surgical/non-surgical) were left to the operator's discretion. An aortography was performed at the end of the procedure for offline evaluation of the regurgitation fraction (RF) by videodensitometry<sup>6,7</sup>. Whenever a post-dilation was performed, a final aortography was performed and analysed with videodensitometry. The quantitative cutoff of RF% on videodensitometry for moderate-severe regurgitation is ≥17%. Postprocedural ECG, echocardiography, and laboratory investigations were performed. Antithrombotic

## Abbreviations

<b>AS</b>	aortic stenosis	<b>MSCT</b>	multislice computed tomography	<b>SEV</b>	self-expanding valve
<b>BEV</b>	balloon-expandable valve	<b>PPI</b>	permanent pacemaker implantation	<b>TAVI</b>	transcatheter aortic valve implantation
<b>ECG</b>	electrocardiography	<b>PVR</b>	prosthetic valve regurgitation	<b>THV</b>	transcatheter heart valve
<b>EOA</b>	effective orifice area	<b>RF</b>	regurgitation fraction	<b>VARC</b>	Valve Academic Research Consortium

therapy was recommended according to the European Society of Cardiology (ESC) and European Association for Cardio-Thoracic Surgery (EACTS) guidelines<sup>8</sup>. At the 30-day follow-up, a clinical assessment, ECG, and echocardiography were performed.

## OUTCOMES

The primary combined safety and effectiveness endpoint at 30 days was a composite of all-cause mortality, all stroke, bleeding (VARC types 3 and 4), acute kidney injury (stages 2, 3, and 4), major vascular complications, moderate or severe prosthetic valve regurgitation (PVR), and conduction system disturbances resulting in a new permanent pacemaker implantation (PPI) as per VARC-3<sup>2</sup>. Secondary endpoints were defined as per VARC-3 and specified in the protocol<sup>3,4</sup>; these included the components of the primary endpoint, technical success, device success, and early safety endpoints at 30-day follow-up. The New York Heart Association (NYHA) Functional Class and the 6-minute walk test were used to measure functional improvement, and a 12-item Short Form Survey was used to gauge quality of life.

The primary and secondary endpoints pertaining to technical and device success were evaluated by an independent clinical events committee that was blinded to the randomisation. The Cardiovascular European Research Centre (Paris, France) analysed ECGs, while the CORRIB Core Laboratory (Galway, Ireland) centrally handled echocardiograms and quantitative assessment of aortographic regurgitation.

## STATISTICAL ANALYSIS

Details of the sample size calculation have been published<sup>1,3</sup>. In brief, assuming an event rate of 26.1% and a non-inferiority margin of 10.44%, the sample size of 768 patients was calculated to demonstrate non-inferiority of the Myval THV series to the contemporary THV series (combined SAPIEN and Evolut THV series) with a statistical power of 93% and a 1-sided alpha of 0.05. With this sample size (Myval THV series: n=384, SAPIEN THV series: n=192, Evolut THV series: n=192), the individual comparison of the Myval THV series versus the SAPIEN THV series, and the Myval THV series versus the Evolut THV series, has a statistical power of 80% with a 1-sided alpha of 0.05 to demonstrate non-inferiority of the Myval THV series to the SAPIEN THV series and to the Evolut THV series, respectively. The non-inferiority assessment of the primary endpoint used a 1-sided 95% confidence interval (CI) calculated using the Farrington-Manning test in the intention-to-treat population. For the subsequent superiority analysis and comparison of the itemised primary endpoint, a proportion test was used to compare the difference between the THV types. Continuous variables are summarised using mean±standard deviation (SD) and median (interquartile range [IQR]) according to distribution and were compared using the 2-sample t-test. Categorical variables are presented as frequency (percentage) and were compared using Pearson's  $\chi^2$  test or Fisher's exact test, as appropriate. The mean difference and risk ratio of the two arms are presented with 95% CIs. Statistical analysis was performed using R software, version 4.3.3 (R Foundation for Statistical Computing).

## Results

### BASELINE CHARACTERISTICS

Between 6 January 2021 and 5 December 2023, 768 patients with severe symptomatic native AS were enrolled, with 384 participants randomly assigned to the Myval THV series, 192 to the SAPIEN THV series, and 192 to the Evolut THV series (**Central illustration**). The consort flow diagram is shown in **Figure 1**. Baseline characteristics, which are tabulated in **Table 1**, were similar between all three arms. The ages, given as mean±SD, were 80.0±5.7, 81.1±5.4, and 79.7±5.4 years in the Myval, SAPIEN and Evolut THV series arms, respectively. The median (IQR) Society of Thoracic Surgeons score was 2.6% (1.7-4.0) in the Myval THV series arm, 2.6% (1.8-4.0) in the SAPIEN THV series arm and 2.7% (1.5-4.0) in the Evolut THV series arm, indicating on average that all arms included a low-risk population. Of note, all three arms included patients with bicuspid valves (6.0% vs 7.3% vs 7.8%) and small aortic annuli (≤430 mm<sup>2</sup>; 32.6% vs 33.3% vs 29.2%). Other baseline demographic characteristics were similar in all three arms.

### PROCEDURAL CHARACTERISTICS

The procedural characteristics are tabulated in **Table 2**.


The mean annular areas were comparable: 470.5±80.0 mm<sup>2</sup>, 469.3±82.6 mm<sup>2</sup> and 473.5±74.2 mm<sup>2</sup> in the Myval, SAPIEN and Evolut THV series arms, respectively. Similarly, the mean annular perimeters were 77.8±6.7 mm, 77.7±6.9 mm and 78.1±6.1 mm, respectively. Owing to crossovers, 379 Myval THV series, 189 SAPIEN THV series and 188 Evolut THV series were implanted. In the Myval THV series arm, the Myval THV (91.3%) was the predominant device, followed by the Myval Octacor (8.7%). In the SAPIEN THV series arm, the SAPIEN 3 (55.4%) and the SAPIEN 3 Ultra (44.6%) were implanted, whilst in the Evolut THV series arm, the most implanted device was Evolut PRO (55.2%) followed by Evolut R (37.0%), Evolut PRO+ (5.2%) and Evolut FX (2.6%). All of the patients in the Evolut THV series arm received 26 mm or above devices; no patient received a 23 mm device. The intermediate-size Myval THV series (21.5 mm, 24.5 mm, and 27.5 mm) constituted 48% of the implanted Myvals; these sizes are not available in the SAPIEN or Evolut THV series. Predilation was performed more frequently in the Myval THV series arm as compared to the SAPIEN THV series arm (43.3% vs 30.7%; p=0.005) and equally as compared to the Evolut THV series arm (43.3% vs 45.7%; p=0.64). Post-dilation rates were comparable between the Myval and SAPIEN THV series arms (10.0% vs 10.1%; p=1.00) and were significantly lower with the Myval THV series as compared to the Evolut THV series (10.0% vs 32.5%; p<0.0001).

### PRIMARY OUTCOME

The probability distribution (with point estimate and 1-sided 95% CI) of the risk difference for the frequency of the primary endpoint between the Myval versus SAPIEN or Evolut THV series arm is depicted in **Figure 2**. The itemised primary outcomes of all three arms are shown in **Figure 3** and **Table 3**.

At 30 days, the primary composite endpoint (non-inferiority analysis) occurred in 24.7% in the Myval THV

## Comparison of clinical and echocardiographic outcomes: Myval vs SAPIEN THV series and Myval vs Evolut THV series.



Niels van Royen *et al.* • EuroIntervention 2024;20:e1-e14 • DOI: 10.4244/EIJ-D-24-00951

A) The LANDMARK trial, a randomised trial comparing the Myval THV series with contemporary THV series (SAPIEN and Evolut) in patients with severe aortic stenosis. B) Comparison between the Myval and SAPIEN THV series. C) Comparison between the Myval and Evolut THV series. THV: transcatheter heart valve

series arm, 24.1% in the SAPIEN THV series arm (absolute risk difference: 0.6%, with the 1-sided upper 95% CI limit of 8.0%) and 30.0% in the Evolut THV series arm (absolute risk difference: -5.3%, with the 1-sided upper 95% CI limit of 2.5%). Therefore, as the predefined non-inferiority margin was 10.44%, the Myval THV series achieved statistically significant non-inferiority compared with the SAPIEN THV series ( $p_{\text{non-inferiority}} = 0.0033$ ) and with the Evolut THV series ( $p_{\text{non-inferiority}} < 0.0001$ ). Secondary analyses of the components of the primary endpoint showed no significant differences between the Myval versus SAPIEN THV series arms or Myval versus Evolut THV series arms (Table 3). Of note, the p-value for the risk difference between the SAPIEN and Myval THV series for bleeding was 0.07 in favour of the SAPIEN THV series (0.5% vs 2.9%) and between the Myval and Evolut THV series, PVR was 0.06 in favour of the Myval THV series (3.4% vs 7.4%). There were no other trends in risk difference among types of valve for events

such as mortality, stroke, PPI, acute kidney injury or major vascular complications.

## SECONDARY OUTCOMES


### TECHNICAL SUCCESS AND DEVICE SUCCESS

Technical and device success rates are shown in **Supplementary Table 1**. Technical success rates at the end of the procedure were 96.3%, 98.9%, and 94.7% in the Myval, SAPIEN, and Evolut THV series arms, respectively. At 30-day follow-up, the device success rates were 91.0%, 92.6%, and 86.7%, respectively.

### CONDUCTION DISTURBANCES AND PPI RATES

The rates of PPI in the Myval, SAPIEN and Evolut THV series arms were 15.0%, 17.3% and 16.8%, respectively, with the underlying indications reported in **Supplementary Table 2**. The new-onset left bundle branch block (LBBB) rates were comparable (Myval THV series: 11.5% [n=39/339], SAPIEN





**Figure 1.** Consort flow diagram of the LANDMARK trial. \*Two patients died before the procedure and were included in the population for endpoint analysis. §Three patients who withdrew consent before the procedure without any known events at that point in time were excluded from the endpoint analysis. \*By Abbott. \*One patient signed informed consent and was randomised; however, the investigators were unaware that the patient had died in the meantime, and therefore the patient was included in the endpoint analysis. §One patient died before the procedure and was included in the population for endpoint analysis. \*\*One patient was excluded after randomisation by the investigator due to rapid progression of his Alzheimer's disease. \*Two participants who withdrew consent before the procedure without any known events at that point in time were excluded from the endpoint analysis. CT: computed tomography; ITT: intention-to-treat; PP: per-protocol; TAVI: transcatheter aortic valve implantation

THV series: 10.0% [n=17/170], Evolut THV series: 14.3% [n=23/161]);  $p_{\text{Myval-SAPIEN}}=0.72$  and  $p_{\text{Myval-Evolut}}=0.46$ .

#### HAEMODYNAMIC PARAMETERS

Echocardiographic assessment of the three arms are shown in **Table 4**. Rates of severe patient-prosthesis mismatch,

based on body mass index, were comparable in the Myval, SAPIEN and Evolut THV series arms at 4.0% (n=15/372), 5.9% (n=11/188) and 1.7% (n=3/179), respectively ( $p_{\text{Myval-SAPIEN}}=0.45$  and  $p_{\text{Myval-Evolut}}=0.23$ ). The rates of moderate-severe PVR were also similar (3.4% vs 1.6% vs 7.4%) (**Table 3**).

**Table 1. Baseline characteristics, medical history and cardiac history of all three cohorts in the LANDMARK trial.**

Characteristics	Myval <sup>a</sup> THV series (n=384)	SAPIEN <sup>b</sup> THV series (n=192)	Evolut <sup>c</sup> THV series (n=192)
Age, years	80.0±5.7	81.1±5.4	79.7±5.4
Sex			
Male	191 (49.7)	106 (55.2)	102 (53.1)
Female	193 (50.3)	86 (44.8)	90 (46.9)
Body mass index, kg/m <sup>2</sup>	28.2±4.9	27.9±4.4	28.2±5.3
Society of Thoracic Surgeons (STS) score, %	2.6 [1.7-4.0]	2.6 [1.8-4.0]	2.7 [1.5-4.0]
Risk category according to STS score			
Low risk (<4%)	290 (75.5)	144 (75.0)	145 (75.5)
Intermediate risk (4-8%)	78 (20.3)	39 (20.3)	39 (20.3)
High risk (>8%)	16 (4.2)	9 (4.7)	8 (4.2)
Estimated glomerular filtration rate <60 ml/min	171/362 (47.2)	85/180 (47.2)	91/180 (50.6)
Estimated glomerular filtration rate <30 ml/min	53/362 (14.6)	31/180 (17.2)	23/180 (12.8)
Small annulus: aortic annulus area ≤430 mm <sup>2</sup>	125 (32.6)	64 (33.3)	56 (29.2)
Bicuspid valve	23 (6.0)	14 (7.3)	15 (7.8)
Diabetes	111 (28.9)	56 (29.2)	58 (30.2)
Hypercholesterolaemia	42 (10.9)	3 (1.6)	33 (17.2)
Hypertension	256 (66.7)	129 (67.2)	125 (65.1)
Chronic obstructive pulmonary disease	42 (10.9)	20 (10.4)	20 (10.4)
History of atrial fibrillation or flutter	94 (24.5)	45 (23.4)	54 (28.1)
Previous stroke	13 (3.4)	3 (1.6)	5 (2.6)
Permanent pacemaker	11 (2.9)	6 (3.1)	12 (6.3)
Previous MI	26 (6.8)	12 (6.3)	11 (5.7)
Previous CABG	13 (3.4)	10 (5.2)	11 (5.7)
History of percutaneous coronary intervention	30 (7.8)	9 (4.7)	16 (8.3)
History of cerebrovascular accident or a transient ischaemic attack in previous 6 months	5 (1.3)	0 (0)	1 (0.5)

Data are presented as n (%), n/N (%), mean±standard deviation, or median [interquartile range]. <sup>a</sup>By Meril Life Sciences; <sup>b</sup>by Edwards Lifesciences; <sup>c</sup>by Medtronic. CABG: coronary artery bypass graft; MI: myocardial infarction; THV: transcatheter heart valve

At 30 days, the Myval THV series had a significantly lower aortic valve mean pressure gradient (MPG) ( $8.2\pm 3.5$  mmHg vs  $10.2\pm 4.9$  mmHg;  $p<0.0001$ ) and higher effective orifice area (EOA) ( $2.02\pm 0.55$  cm<sup>2</sup> vs  $1.80\pm 0.52$  cm<sup>2</sup>;  $p<0.0001$ ) compared to the SAPIEN THV series (Table 4), whereas it had a significantly higher MPG ( $8.2\pm 3.5$  mmHg vs  $5.6\pm 2.3$  mmHg;  $p<0.0001$ ) and lower EOA ( $2.02\pm 0.55$  cm<sup>2</sup> vs  $2.31\pm 0.55$  cm<sup>2</sup>;  $p<0.0001$ ) compared to the Evolut THV series (Table 4). A THV size-specific comparison of the mean EOA of the 23 mm ( $1.80\pm 0.49$  cm<sup>2</sup> vs  $1.58\pm 0.49$  cm<sup>2</sup>;  $p=0.01$ ), 26 mm ( $2.13\pm 0.52$  cm<sup>2</sup> vs  $1.90\pm 0.45$  cm<sup>2</sup>;  $p=0.0041$ ) and 29 mm ( $2.43\pm 0.61$  cm<sup>2</sup> vs  $2.05\pm 0.48$  cm<sup>2</sup>;  $p=0.01$ ) Myval and SAPIEN valves showed that the mean EOA with the Myval THV series was significantly larger than with the SAPIEN THV series. The mean EOAs of the 20 mm nominal sizes of the Myval and SAPIEN THV series ( $1.42\pm 0.05$  cm<sup>2</sup> vs  $1.38\pm 0.36$  cm<sup>2</sup>;  $p=0.78$ ) were comparable (Figure 4). A similar THV size-specific comparison of the mean EOAs of the 26 mm ( $2.13\pm 0.52$  cm<sup>2</sup> vs  $2.22\pm 0.44$  cm<sup>2</sup>;  $p=1.00$ ) and 29 mm ( $2.43\pm 0.61$  cm<sup>2</sup> vs  $2.27\pm 0.52$  cm<sup>2</sup>;  $p=0.48$ ) Myval and Evolut THV series showed no significant difference (Figure 4). None of the patients were implanted

with a 23 mm Evolut valve and, per design and protocol, the 30.5 mm and 32 mm Myval THV series were not included in the randomised trial and thus not available for comparison with the 34 mm Evolut THV series.

In terms of RF% assessed by quantitative aortography on the final angiogram, the RF (median 3.0% [1<sup>st</sup>, 3<sup>rd</sup> quartiles: 1.0, 7.0]) was comparable between the Myval and SAPIEN THV series, whilst the difference between the Myval and Evolut THV series (3.0% [1.0, 7.0] and 5.0% [1.0, 10.0];  $p=0.0007$ ) was highly significant. An RF higher than 17% was documented in 2%, 4% and 8% of the patients in the Myval, SAPIEN ( $p_{\text{Myval-SAPIEN}}=0.23$ ) and Evolut ( $p_{\text{Myval-Evolut}}=0.0057$ ) THV series arms, respectively (Table 2).

#### QUALITY OF LIFE


In all three arms, there were significant ( $p<0.0001$ ) improvements in the NYHA Functional Class (Supplementary Figure 1), distance covered in the six-minute walk test (Supplementary Table 3), and physical and mental quality-of-life scores (Supplementary Table 4) between baseline and 30-day follow-up; however, no differences were noted between the Myval versus SAPIEN THV series and the Myval versus Evolut THV series.

**Table 2. Procedural characteristics of all three cohorts in the LANDMARK trial.**

Procedural details	Myval THV series n=384	SAPIEN THV series n=192	Evolut THV series n=192	p-value Myval vs SAPIEN	p-value Myval vs Evolut
<b>Access site</b>					
Transfemoral	378 (99.7)	188 (97.5)	188 (100.0)	1.00	1.00
Right femoral	334 (88.1)	167 (88.4)	168 (89.4)	0.98	0.83
Percutaneous	297 (78.4)	158 (83.6)	141 (75.0)	0.06	0.15
Surgical cutdown	37 (9.8)	9 (4.8)	27 (14.4)		
Left femoral	44 (11.6)	21 (11.1)	20 (10.6)	0.98	0.83
Percutaneous	40 (10.6)	20 (10.6)	18 (9.6)	1.00	1.00
Surgical cutdown	4 (1.1)	1 (0.5)	2 (1.1)		
Subclavian	1 (0.3)	1 (0.5)	0 (0)	1.00	1.00
Right subclavian artery	0 (0)	1 (0.5)	0 (0)	1.00	1.00
Percutaneous	0 (0)	0 (0)	0 (0)	1.00	1.00
Surgical cutdown	0 (0)	1 (0.5)	0 (0)		
Left subclavian artery	1 (0.3)	0 (0)	0 (0)	1.00	1.00
Percutaneous	0 (0)	0 (0)	0 (0)	1.00	1.00
Surgical cutdown	1 (0.3)	0 (0)	0 (0)		
Annular area, mm <sup>2</sup>	470.5±80.0 (n=384)	469.3±82.6 (n=192)	473.5±74.2 (n=192)	0.81	0.60
Annular perimeter, mm	77.8±6.7 (n=384)	77.7±6.9 (n=192)	78.1±6.1 (n=192)	0.72	0.60
Procedural time, min	77.0±40.3 (n=378)	76.5±43.2 (n=189)	78.7±37.1 (n=188)	0.63	0.31
Contrast volume, ml	143.6±68.5 (n=355)	144.9±65.6 (n=189)	155.2±79.1 (n=175)	0.64	0.21
General anaesthesia	73 (19.3) (n=379)	24 (12.7) (n=189)	50 (26.6) (n=188)	0.07	0.06
Conscious sedation	306 (80.7) (n=379)	165 (87.3) (n=189)	138 (73.4) (n=188)		
Predilation	164 (43.3) (n=379)	58 (30.7) (n=189)	86 (45.7) (n=188)	0.005	0.64
TAVI device implanted	379	189	188		
RF after implantation, prior to post-dilation, %	12.0 (6.0, 18.5) (n=23)	18.0 (1.0, 19.0) (n=9)	10.5 (6.0, 15.0) (n=26)	0.79	0.62
RF >17% after implantation, prior to post-dilation	6 (26.1) (n=23)	5 (55.6) (n=9)	6 (23.1) (n=26)	0.21	1.00
Post-dilation	38 (10.0) (n=379)	19 (10.1) (n=189)	61 (32.5) (n=188)	1.00	<0.0001
RF after post-dilation, %	2.0 (1.0, 8.0) (n=33)	3.0 (2.0, 8.0) (n=17)	5.0 (1.0, 9.5) (n=47)	0.37	0.21
RF >17% after post-dilation	0 (0) (n=33)	1 (5.9) (n=17)	4 (8.5) (n=47)	0.34	0.14
RF in final aortogram, %	3.0 (1.0, 7.0) (n=295)	3.0 (1.0, 7.0) (n=151)	5.0 (1.0, 10.0) (n=150)	0.86	0.0007
RF >17%	6 (2.0) (n=295)	6 (4.0) (n=151)	12 (8.0) (n=150)	0.23	0.006
Cerebral protection device	48 (13.2)	12 (6.8)	21 (11.4)	0.03	0.71
Use of closure device	344 (90.8)	180 (95.2)	160 (85.1)	0.09	0.06
Length of hospital stay, days	4.0 (3.0, 6.0) (n=374)	4.0 (2.0, 6.0) (n=189)	4.0 (2.0, 6.0) (n=186)	0.72	0.66

Data are presented as n, n (%), mean±standard deviation, or median (Q1, Q3). Q: quartile; RF: regurgitation fraction assessed by videodensitometry of the aortography; TAVI: transcatheter aortic valve implantation; THV: transcatheter heart valve





## PATIENTS WITH A SMALL AORTIC ANNULUS

The clinical and echocardiographic outcomes in patients with a small aortic annulus ( $\leq 430 \text{ mm}^2$ ) are shown in **Supplementary Table 5** and **Table 5**. The event rates of the primary composite endpoint were comparable between the three arms (Myval THV series: 20%, SAPIEN THV series: 21% and Evolut THV series: 33%;  $p_{\text{Myval-SAPIEN}}=1.00$  and  $p_{\text{Myval-Evolut}}=0.08$ ) (**Supplementary Table 5**). The mean EOA was significantly larger in the Evolut THV series ( $2.27 \pm 0.49 \text{ cm}^2$ ) arm than in the Myval THV series arm ( $1.75 \pm 0.49 \text{ cm}^2$ ) or SAPIEN THV series arm ( $1.53 \pm 0.45 \text{ cm}^2$ ) ( $p_{\text{Myval-SAPIEN}}=0.006$  and  $p_{\text{Myval-Evolut}} < 0.0001$ ). The MPG was significantly lower in the Myval THV series arm compared to the SAPIEN THV series arm ( $9.30 \pm 3.74 \text{ mmHg}$  vs  $11.78 \pm 5.40 \text{ mmHg}$ ;  $p=0.0005$ ), and the Evolut THV series arm had a significantly lower MPG than the Myval THV series arm ( $5.76 \pm 2.33 \text{ mmHg}$ ;  $p < 0.0001$ ) (**Table 5**).

## Discussion

### COMPOSITE CLINICAL PRIMARY ENDPOINTS AMONG THE THREE VALVES

This substudy of the LANDMARK trial compared outcomes for the first time between two BEVs – the Myval


THV series and the SAPIEN THV series – in addition to comparing the Myval THV series to the SE Evolut THV series. The key findings of this substudy are the individual non-inferiority of the Myval THV series to the SAPIEN THV series and to the Evolut THV series for the primary composite safety and effectiveness endpoint at 30-day follow-up. Additionally, no significant differences were found between the arms for any components of the primary composite endpoint.

### HAEMODYNAMIC ASSESSMENT AMONG THE THREE VALVES

At 30 days, the MPG ( $p < 0.0001$ ) and EOA ( $p < 0.0001$ ) were significantly better with the Myval THV series than the SAPIEN THV series and with the Evolut THV series than the Myval THV series.

Notably, there were no significant differences in EOA between the Myval and Evolut THV series for the 26 mm ( $2.13 \pm 0.52 \text{ cm}^2$  vs  $2.22 \pm 0.44 \text{ cm}^2$ ;  $p=1.00$ ) or 29 mm ( $2.43 \pm 0.61 \text{ cm}^2$  vs  $2.27 \pm 0.52 \text{ cm}^2$ ;  $p=0.48$ ) diameter valves (**Figure 4**). The overall better haemodynamics of the SE Evolut THV series in the LANDMARK trial is mainly due to the more favourable EOA and gradients observed in the patients with a small annulus who were exclusively treated with a 26 mm THV, instead of a 23 mm THV. These findings align with the SMART trial, which also used a minimal (2.3%) number of 23 mm Evolut THVs<sup>9</sup>. As a matter of fact, EOA and gradients in all prior studies have been reported based on echocardiographic assessment. Although echocardiographic assessments generally report better haemodynamics for supra-annular SE valves (SEVs), controversy remains about their durability compared to BEVs, suggesting a need for invasive gradient analysis<sup>10</sup>. Only one case in the Myval THV series arm received a 30.5 mm THV (protocol deviation) with a resultant EOA of  $3.30 \text{ cm}^2$ , whereas in the Evolut 34 mm group ( $n=44$ ), the average EOA was  $2.42 \pm 0.65 \text{ cm}^2$ , similar to the Myval 29 mm group ( $2.43 \pm 0.61 \text{ cm}^2$ ;  $p=0.95$ ).

In patients with a small aortic annulus, the mean EOAs were significantly larger in the Evolut THV series arm than in the Myval and SAPIEN THV series arms. In a recent meta-analysis of 21 studies ( $n=8,647$ ) comparing SEVs and BEVs in small aortic annuli, SEVs had superior haemodynamics, but higher rates of paravalvular leak, PPI and in-hospital stroke<sup>11</sup>. The significantly larger EOA of the Evolut THV series in patients with a small aortic annulus may represent a drawback in the LANDMARK trial: there were seven instances of PVR (7/59, 11.9%) in the Evolut THV series arm compared to two (2/117, 1.7%;  $p=0.007$ ) in the Myval THV series arm (**Supplementary Figure 2**). The observed relatively high PVR rate with the Evolut THV series in small annulus patients may be related to non-uniform expansion of the SEV. Moscarelli et al reported that non-uniform expansion is consistently observed after implantation of a SEV, with eccentricity more frequent at the annular level compared to the prosthesis frame outflow level<sup>12</sup>. It has also been demonstrated that underexpansion and non-uniform expansion of the SEV could result in an elliptical shape of the stent frame at the level of leaflet coaptation, which is associated with an increased incidence of PVR and putatively resulted in a pinwheeling effect that



could generate strain on the leaflet and affect a THV's long-term durability<sup>13</sup>.

The overall comparison of moderate/severe PVR between the Myval and Evolut (including Evolut R) THV series showed a p-value of 0.06 (Myval THV series: 3.4% vs Evolut THV series: 7.4%; risk difference: -4.0%, 95% CI: -8.5 to 0.6). However, a sensitivity analysis for moderate-severe PVR between the Myval and Evolut THV series, after excluding the Evolut R, showed a p-value of 0.28 (Myval THV series: 3.4% vs Evolut THV series: 6.1%; risk difference: -2.68%, 95% CI: -7.98 to 2.63) (**Supplementary Table 6**).

The incidence of moderate-severe PVR in the LANDMARK trial is similar to in the SCOPE I and SOLVE-TAVI trials, showing that BEVs have lower rates than SEVs<sup>14,15</sup>. Balloon post-dilation (BPD) is a commonly used technique for minimising the degree of PVR following TAVI<sup>16</sup>. However, it is associated with serious complications such as annular

rupture, stroke, and damage to the prosthetic leaflets, which may increase the risk of early THV deterioration<sup>17</sup>. A significantly lower proportion of patients required BPD in the BE Myval and SAPIEN THV series arms (10.0% and 10.1%) compared to the SE Evolut THV series arm (32.5%; p<0.0001).

Notably, contemporary THV designs with better sealing skirts have gradually reduced the frequency of more-than-mild PVR<sup>18,19</sup>. The lower rates of PVR with the Myval THV series could be due to the internal and external skirt design (reduces PVR) and the availability of intermediate sizes, which eliminates the need for over- and undersizing and results in an ideal fit to the native annulus.

A key innovation of the Myval THV series is its availability of intermediate sizes with 1.5 mm differences, compared to the conventional 3 mm step-up in nominal sizes. Our study found that about half of the Myval

**Table 3. Primary outcomes of all three cohorts in the LANDMARK trial.**


Events	Myval THV series n=384	SAPIEN THV series n=192	Evolut THV valves n=192	Risk difference* Myval vs SAPIEN	p-value Myval vs SAPIEN	Risk difference* Myval vs Evolut	p-value Myval vs Evolut
<b>Primary analysis</b>							
Primary endpoint (non-inferiority analysis)	94/381 (24.7)	46/191 (24.1)	57/190 (30.0)	0.6 (NA to 8.0)	0.0033	-5.3 (NA to 2.5)	<0.0001
<b>Secondary analysis</b>							
Primary endpoint (superiority analysis)	94/381 (24.7)	46/191 (24.1)	57/190 (30.0)	0.6 (-7.3 to 8.4)	0.96	-5.3 (-13.6 to 2.9)	0.21
<b>Individual components of the primary endpoint</b>							
All-cause mortality	9/381 (2.4)	3/191 (1.6)	6/190 (3.2)	0.8 (-1.9 to 3.5)	0.76	-0.8 (-4.1 to 2.5)	0.59
All stroke	12/381 (3.1)	6/191 (3.1)	6/190 (3.2)	0.0 (-3.0 to 3.0)	1.00	-0.1 (-3.1 to 3.0)	1.00
Bleeding (type 3 and type 4)	11/381 (2.9)	1/191 (0.5)	4/190 (2.1)	2.4 (0.0 to 4.7)	0.07	0.8 (-2.3 to 3.8)	0.78
Acute kidney injury (stage 2, stage 3 and stage 4)	6/381 (1.6)	0/191 (0)	3/190 (1.6)	1.6 (-0.1 to 3.2)	0.19	0.0 (-2.2 to 2.2)	1.00
Moderate or severe prosthetic valve regurgitation	13/381 (3.4)	3/191 (1.6)	14/190 (7.4)	1.8 (-1.1 to 4.8)	0.32	-4.0 (-8.5 to 0.6)	0.06
Conduction system disturbances resulting in a new permanent pacemaker	57/381 (15.0)	33/191 (17.3)	32/190 (16.8)	-2.3 (-9.2 to 4.5)	0.55	-1.8 (-8.7 to 4.9)	0.64
Major vascular complications	6/381 (1.6)	2/191 (1.0)	7/190 (3.7)	0.6 (-1.8 to 2.8)	0.72	-2.1 (-5.5 to 1.2)	0.14

Data are n/N (%) or risk difference (95% confidence interval). \*All 95% confidence intervals and p-values are two-sided except those of the primary composite endpoint analysis for non-inferiority (one-sided). THV: transcatheter heart valve

**Table 4. Echocardiographic data for Myval versus SAPIEN THV series and Myval versus Evolut THV series.**

<b>Myval vs SAPIEN THV series</b>											
Parameter	Baseline			Discharge			30 days			p-value (baseline vs 30 days)	
	Myval THV series	SAPIEN THV series	p-value	Myval THV series	SAPIEN THV series	p-value	Myval THV series	SAPIEN THV series	p-value	Myval THV series	SAPIEN THV series
Effective orifice area, cm <sup>2</sup>	0.74±0.22 (n=364)	0.70±0.21 (n=180)	0.04	2.16±0.61 (n=353)	1.84±0.51 (n=175)	<0.0001	2.02±0.55 (n=346)	1.80±0.52 (n=169)	<0.0001	<0.0001	<0.0001
AV mean pressure gradient, mmHg	39.9±14.0 (n=368)	39.2±14.2 (n=184)	0.57	8.3±4.0 (n=362)	10.9±4.6 (n=181)	<0.0001	8.2±3.5 (n=355)	10.2±4.9 (n=174)	<0.0001	<0.0001	<0.0001
Aortic regurgitation assessment	n=360	n=186		n=362	n=184		n=350	n=171			
None/trace-mild	318 (88.3)	165 (88.7)	1.00	351 (97.0)	181 (98.4)	0.40	341 (97.4)	168 (98.3)	0.76	<0.0001	0.0007
Moderate-severe	42 (11.7)	21 (11.3)	1.00	11 (3.0)	3 (1.6)	0.40	9 (2.6)	3 (1.8)	0.76	<0.0001	0.0007
<b>Myval vs Evolut THV series</b>											
Parameter	Baseline			Discharge			30 days			p-value (baseline vs 30 days)	
	Myval THV series	Evolut THV series	p-value	Myval THV series	Evolut THV series	p-value	Myval THV series	Evolut THV series	p-value	Myval THV series	Evolut THV series
Effective orifice area, cm <sup>2</sup>	0.74±0.22 (n=364)	0.74±0.23 (n=180)	0.91	2.16±0.61 (n=353)	2.35±0.56 (n=171)	0.0003	2.02±0.55 (n=346)	2.31±0.55 (n=168)	<0.0001	<0.0001	<0.0001
AV mean pressure gradient, mmHg	39.9±14.0 (n=368)	38.2±12.9 (n=184)	0.16	8.3±4.0 (n=362)	5.9±2.5 (n=175)	<0.0001	8.2±3.5 (n=355)	5.6±2.3 (n=175)	<0.0001	<0.0001	<0.0001
Aortic regurgitation assessment	n=360	n=182		n=362	n=173		n=350	n=174			
None/trace-mild	318 (88.3)	156 (85.7)	0.46	351 (97.0)	161 (93.1)	0.06	341 (97.4)	163 (93.7)	0.06	<0.0001	0.02
Moderate-severe	42 (11.7)	26 (14.3)	0.46	11 (3.0)	12 (6.9)	0.06	9 (2.6)	11 (6.3)	0.06	<0.0001	0.02

Data are presented as n (%) or mean±standard deviation. AV: aortic valve; THV: transcatheter heart valve



**Figure 4.** Scatter plot of postprocedural EOA (cm<sup>2</sup>) as assessed by echocardiography, and preprocedural aortic annulus area (mm<sup>2</sup>) as assessed by computed tomography, categorised according to the nominal size of the three different valve series. \*P-values are based on 2-sample t-test. EOA: effective orifice area

patients were implanted with these intermediate sizes based on preprocedural MSCT assessments, potentially preventing oversizing or undersizing. The Myval THV series shows a higher and narrower density curve of fitting index compared to the SAPIEN THV series (**Supplementary Figure 3**). More patients implanted with a Myval THV series had a fitting index (the ratio between the nominal THV diameter and MSCT-derived aortic annulus diameter) around 1.0, indicating proper fit, which can be attributed to the wider range of sizes. This better sizing and fitting may

contribute to the superior EOA and lower transvalvular gradient in the Myval THV series (**Supplementary Table 7**) and may help reduce the occurrence of PVR and PPI<sup>20</sup>. With appropriate fitting, PPI rates were similar between the two valves (15.0% vs 14.1%), but there was a numerical difference favouring the Myval THV series (15.2% vs 21.1%) when fitting was more appropriate, possibly due to the use of intermediate sizes. A future pooled analysis of the LANDMARK and Compare-TAVI trials may validate these hypotheses.

#### OCCURRENCE OF PPI AMONG THE THREE VALVES

New PPI rates were comparable among the three arms (Myval THV series: 15.0%, SAPIEN THV series: 17.3%, Evolut THV series: 16.8%). Previous reports found that SEV implantation is a predictor for new PPI after TAVI, next to age, baseline right bundle branch block (RBBB), baseline LBBB and THV implantation depth<sup>21</sup>.

A recent meta-analysis of 23 studies (n=18,610) reported a crude incidence of 17% for PPI (range 8.8% to 32%), consistent with our results. However, SEVs and baseline RBBB were associated with a 2-fold greater risk of continued pacemaker dependency 1 year after TAVI<sup>22</sup>.

The cusp-overlap view technique for SEVs and a high deployment technique for BEVs and SEVs have proven helpful in reducing conduction abnormalities and new PPI<sup>23,24</sup>. The Myval THV series shows less shortening (Myval Octacore: 19-20%, Myval THV series: 21-24%, SAPIEN THV series: 26-27%, Evolut THV series: 44%), facilitates precision placement and deployment accuracy, and may reduce PVR and PPI requirement<sup>25</sup>. Again, these remain hypotheses until outcomes from the pooled analysis (>1,500 patients) from LANDMARK and Compare-TAVI are available.

Technological improvements with new THV devices and overcoming the learning curve could also help to reduce the need for PPI. It is crucial to analyse each baseline ECG for future PPI risk and to assess the patient's anatomy before

**Table 5.** Echocardiographic data in patients with a small aortic annulus (≤430 mm<sup>2</sup>).

Parameters	Baseline			Discharge			30-day follow-up			p-value (at 30 days)	
	Myval THV series	SAPIEN THV series	Evolut THV series	Myval THV series	SAPIEN THV series	Evolut THV series	Myval THV series	SAPIEN THV series	Evolut THV series	Myval vs SAPIEN	Myval vs Evolut
Effective orifice area, cm <sup>2</sup>	0.70±0.20 (n=118)	0.65±0.19 (n=60)	0.72±0.24 (n=52)	1.86±0.55 (n=113)	1.62±0.41 (n=60)	2.28±0.52 (n=48)	1.75±0.49 (n=112)	1.53±0.45 (n=58)	2.27±0.49 (n=47)	0.006	<0.0001
AV mean pressure gradient, mmHg	41.05±13.52 (n=119)	42.53±14.92 (n=60)	39.88±14.37 (n=54)	9.48±4.88 (n=116)	12.73±4.89 (n=62)	6.14±2.76 (n=49)	9.30±3.74 (n=116)	11.78±5.40 (n=59)	5.76±2.33 (n=50)	0.0005	<0.0001
Aortic regurgitation grade	n=112	n=61	n=53	n=115	n=61	n=51	n=109	n=58	n=52	-	-
None or trace	39 (35)	18 (30)	14 (26)	81 (70)	52 (85)	20 (39)	74 (68)	47 (81)	23 (44)	0.03	0.0003
Mild	58 (52)	28 (46)	31 (58)	31 (27)	8 (13)	24 (47)	35 (32)	10 (17)	23 (44)		
Moderate	9 (8)	14 (23)	8 (15)	1 (1)	0 (0)	5 (10)	0 (0)	0 (0)	5 (10)		
Severe	2 (2)	0 (0)	0 (0)	0 (0)	0 (0)	0 (0)	0 (0)	0 (0)	0 (0)		
Not evaluable	4 (4)	1 (2)	0 (0)	2 (2)	1 (2)	2 (4)	0 (0)	1 (2)	1 (2)		

Data are presented as mean±standard deviation or n (%). AV: aortic valve; THV: transcatheter heart valve



TAVI, as patients with a shorter membranous septum may benefit from the Minimizing Depth According to the membranous Septum (MIDAS) technique<sup>26</sup>. However, it is equally important to recognise that, in 23% of patients, the membranous septum terminates above the annular plane, leaving the operator no room to manoeuvre<sup>27</sup>.

## Limitations

Our study has a few limitations. First, multiple THV iterations were used across all arms, limiting representation of the latest devices, which were unavailable in Europe during enrolment (e.g., SAPIEN Resilia). Second, the decision for new PPI was left to the investigator's discretion, potentially introducing bias. Third, the study evaluated only early 30-day outcomes; long-term outcomes are essential for robust device comparison. For the following 10 years, clinical and echocardiographic evaluations will be carried out together with ongoing data monitoring for long-term analyses. While the current findings are encouraging, they must be verified in long-term follow-up. The superiority in haemodynamic performance of SEVs compared to BEVs in small aortic annulus patients needs further investigation to determine if these short-term benefits are sustained clinically, haemodynamically, and in terms of durability.

## Conclusions

In conclusion, this prespecified substudy of the LANDMARK trial demonstrated that the Myval THV series was non-inferior to the SAPIEN THV series and the Evolut THV series in terms of early safety and effectiveness at 30 days in elderly patients with severe, symptomatic aortic stenosis.

## Authors' affiliations

1. Department of Cardiology, Radboud University Hospital, Nijmegen, the Netherlands; 2. Centro de Investigación Biomédica en Red - Enfermedades Cardiovasculares (CIBERCV), Instituto de Salud Carlos III, Madrid, Spain; 3. Department of Cardiology, Hospital Clínico Universitario de Valladolid, Valladolid, Spain; 4. Department of Acute Cardiology, Middle-Slovak Institute of Cardiovascular Diseases, Banská Bystrica, Slovakia; 5. Department of Cardiology, University Medical Centre Ljubljana, Ljubljana, Slovenia; 6. Department of Cardiology, Amphia Hospital, Breda, the Netherlands; 7. Department of Interventional Cardiology, Maastricht University Medical Center, Maastricht, the Netherlands; 8. Zuyderland Hospital, Limburg, the Netherlands; 9. Department of Invasive Cardiology, North Estonia Medical Centre, Tallinn, Estonia; 10. Department of Cardiac and Transplant Surgery, University Hospital Dubrava, Zagreb, Croatia; 11. Heart and Vascular Center, Semmelweis University, Budapest, Hungary; 12. Department of Cardiology, Isala Hospital, Zwolle, the Netherlands; 13. Department of Cardiology, European Interbalkan Medical Center, Thessaloniki, Greece; 14. Department of Cardiology, Institute of Heart Diseases, Wrocław Medical University, Wrocław, Poland; 15. Department of Cardiology, St Antonius Hospital, Nieuwegein, the Netherlands; 16. Department of Interventional Cardiology, University Hospital of Gran Canaria Dr Negrín, Las Palmas de Gran Canaria, Spain; 17. Department of Invasive Cardiology, Instituto Dante Pazzanese, São Paulo, Brazil; 18. CHRC, NOVA Medical School, NOVA University

Lisbon, Lisbon, Portugal; 19. Department of Interventional Cardiology, Lille University, Lille, France; 20. Department of Internal Medicine and Cardiology, University Clinic, Heart Center Dresden, University of Technology Dresden, Dresden, Germany; 21. Department of Interventional Cardiology, IRCCS Ospedale Galeazzi Sant'Ambrogio, Milan, Italy; 22. School of Medicine, Vita-Salute San Raffaele University, Milan, Italy; 23. Interventional Cardiology Unit, IRCCS San Raffaele Scientific Institute, Milan, Italy; 24. Department of Cardiology, Auckland City Hospital, Auckland, New Zealand; 25. Department of Cardiology, Hippokration Hospital, Athens, Greece; 26. Department of Interventional Cardiology, Henri Mondor University Hospital, Créteil, France; 27. Department of Clinical Cardiology, San Donato Hospital, Milan, Italy; 28. Department of Cardiology, University Medical Center Utrecht, Utrecht, the Netherlands; 29. Department of Cardiology, University Hospital Reina Sofía, University of Córdoba, IMIBIC, CIBERCV, Córdoba, Spain; 30. Department of Cardiology, Sahlgrenska University Hospital, Gothenburg, Sweden; 31. Department of Molecular and Clinical Medicine, Institute of Medicine, University of Gothenburg, Gothenburg, Sweden; 32. Department of Cardiology & Angiology, University of Giessen and Marburg, Gießen, Germany; 33. Department of Cardiology, Kerckhoff Heart Center, Bad Nauheim, Germany; 34. Department of Cardiology and Angiology, Campus Bad Krozingen, University Heart Center-University of Freiburg, Bad Krozingen, Germany; 35. Faculty of Medicine, University of Freiburg, Freiburg, Germany; 36. Department of Cardiology, University Hospital of Split, Split, Croatia; 37. Department of Cardiology University Hospital Son Espases, Health Research Institute of the Balearic Islands (IdISBa), Palma, Balearic Islands, Spain; 38. Department of Cardiology, School of Medicine, University of Galway, Galway, Ireland; 39. Cardiovascular Center, Taichung Veterans Hospital, Taiwan; 40. Department of Cardiology, Royal Blackburn Hospital, Blackburn, United Kingdom; 41. School of Medicine, University of Central Lancashire, Preston, United Kingdom; 42. Department of Clinical Research, Meril Life Sciences Pvt. Ltd., Vapi, India; 43. Cardiovascular European Research Center (CERC), Massy, France; 44. ICPS, Hôpital privé Jacques Cartier, Massy, France; 45. Galway University Hospital, Galway, Ireland; 46. Centre for Cardiovascular Medicine and Devices, William Harvey Research Institute, Queen Mary University of London and Barts Heart Centre, London, United Kingdom; 47. Cleveland Clinic, London, United Kingdom

## Funding

This trial was funded by Meril Life Sciences. The funder provided all financial support to conduct this trial. The trial protocol, the data analysis and the writing of the manuscript were supervised by the steering committee in consultation with the funder.

## Conflict of interest statement

N. van Royen reports grant funding and personal fees from Abbott; grants from Philips, Biotronik, and Medtronic; speaker fees from MicroPort, Bayer, and RainMed Medical outside the submitted work; and travel support to attend meeting from Meril Life Sciences. I.J. Amat-Santos reports being a proctor for Medtronic, Boston Scientific, and Meril



Life Sciences. A. Ijsselmuiden reports consulting fees from Meril Life Sciences, Angiocare, PulseCath, and Cardiawave; received an institutional grant from Medtronic and Abbott. D. Unic reports payment for workshops from Medtronic; and is a member of the Medtronic EMEA surgical advisory board. B. Merkely reports institutional grants and speaker fees from Abbott, AstraZeneca, Biotronik, Boehringer Ingelheim, CSL Behring, Daiichi Sankyo, DUKE Clinical Institute, Medtronic, and Novartis; institutional fees from Boston Scientific, Bristol-Myers Squibb, Eli Lilly, Terumo, and VIFOR Pharma. R.S. Hermanides reports speaker fees from Amgen, Novartis, Edwards Lifesciences, Meril Life Sciences, and Abbott outside the submitted work. P. Martin reports a proctorship grant from Meril Life Sciences; and an educational event grant from Medtronic. E. van Belle is president of the French Interventional Working Group (GACI) and a Board member of EAPCI. A. Linke received grants from Edwards Lifesciences and Novartis; speaker honoraria from Edwards Lifesciences, Abiomed, Abbott, Boston Scientific, Novartis, Pfizer, BMS, Daiichi Sankyo, AstraZeneca, Boehringer Ingelheim, Meril Life Sciences, and Corvia; travel support from AstraZeneca, Abbott, Meril Life Sciences, Abiomed and Boston Scientific; is a partial patent holder with Boston Scientific; is a stock option holder with Pi-Cardia, Transverse Medical and Filterlex. K. Toutouzas reports proctorship with Abbott, Meril Life Sciences and Medtronic; consulting fees from Gore Medical; is a Board member of the Hellenic Society of Cardiology. M. De Sousa Almeida reports lecture fees from Medtronic and Novartis; travel support from Medtronic, Terumo and Boston Scientific. F. Bedogni reports grants, consulting fees, payment/honoraria/speaker fees and travel support to attend meetings from Medtronic, Abbott, Boston Scientific and Meril Life Sciences; and reports participation in a DSMB for Abbott. M. Pan reports lecture fees from Medtronic and Abbott. O. Angerås reports proctorship and speaker fees with Meril Life Sciences and Abbott; speaker fees from Medtronic; and support for attending meetings from Meril Life Sciences. W.-K. Kim reports honoraria or consultancy fees from Edwards Lifesciences; consulting fees from Boston Scientific, Meril Life Sciences and Abbott; and participation on data and safety monitoring board for HID Imaging and P&F. J. Rothe reports personal fees for consulting/proctoring from Meril Life Sciences, Medtronic, and Qatna; and travel support for attending meetings from Meril Life Sciences, Edwards Lifesciences, Abbott, Medtronic, and Boston Scientific. D. Westermann reports personal fees from Abiomed, AstraZeneca, Boehringer Ingelheim, Novartis, Meril Life Sciences and Medtronic. A. Tobe reports a grant from the Fukuda Foundation for Medical Technology. S. Garg reports honoraria or consultancy fees from Biosensors. U. Chandra and A. Thakkar are full employees of Meril Life Sciences. M.-C. Morice reports that she is shareholder and CEO of CERC, a CRO involved in the trial; and minor shareholder of Electroducer. O. Soliman reports research grants from Biosensors, Boston Scientific, Cardiawave and Meril Life Sciences. P.W. Serruys reports consultancy fees from SMT, Novartis, Meril Life Sciences, and Philips. A. Baumbach reports consultation and speaker fees from AstraZeneca, Sinomed, MicroPort, Medtronic, Faraday, Pi-Cardia, Biosensors, JenaValve and Meril Life Sciences. The other authors have no conflicts of interest to declare.

## References

1. Baumbach A, van Royen N, Amat-Santos IJ, Hudec M, Bunc M, Ijsselmuiden A, Laanmets P, Unic D, Merkely B, Hermanides RS, Ninios V, Protasiewicz M, Rensing BJWM, Martin PL, Feres F, De Sousa Almeida M, van Belle E, Linke A, Ielasi A, Montorfano M, Webster M, Toutouzas K, Teiger E, Bedogni F, Voskuil M, Pan M, Angerås O, Kim WK, Rothe J, Kristić I, Peral V, Garg S, Elzomor H, Tobe A, Morice MC, Onuma Y, Soliman O, Serruys PW; LANDMARK trial investigators. LANDMARK comparison of early outcomes of newer-generation Myval transcatheter heart valve series with contemporary valves (Sapien and Evolut) in real-world individuals with severe symptomatic native aortic stenosis: a randomised non-inferiority trial. *Lancet*. 2024;403:2695-708.
2. VARC-3 WRITING COMMITTEE; Généreux P, Piazza N, Alu MC, Nazif T, Hahn RT, Pibarot P, Bax JJ, Leipsic JA, Blanke P, Blackstone EH, Finn MT, Kapadia S, Linke A, Mack MJ, Makkar R, Mehran R, Popma JJ, Reardon M, Rodes-Cabau J, Van Mieghem NM, Webb JG, Cohen DJ, Leon MB. Valve Academic Research Consortium 3: updated endpoint definitions for aortic valve clinical research. *Eur Heart J*. 2021;42:1825-57.
3. Kawashima H, Soliman O, Wang R, Ono M, Hara H, Gao C, Zeller E, Thakkar A, Tamburino C, Bedogni F, Neumann FJ, Thiele H, Abdel-Wahab M, Morice MC, Webster M, Rosseel L, Mylotte D, Onuma Y, Wijns W, Baumbach A, Serruys PW. Rationale and design of a randomized clinical trial comparing safety and efficacy of myval transcatheter heart valve versus contemporary transcatheter heart valves in patients with severe symptomatic aortic valve stenosis: The LANDMARK trial. *Am Heart J*. 2021;232:23-38.
4. Tobe A, Onuma Y, Soliman O, Baumbach A, Serruys PW. LANDMARK trial: Update in study protocol. *Am Heart J*. 2024;270:162-3.
5. Frane JW. A Method of Biased Coin Randomization, Its Implementation, and Its Validation. *Drug Information Journal*. 1998;32:423-32.
6. Abdelshafy M, Serruys PW, Tsai TY, Revaiah PC, Garg S, Aben JP, Schultz CJ, Abdelghani M, Tonino PAL, Miyazaki Y, Rutten MCM, Cox M, Sahyoun C, Teng J, Tateishi H, Abdel-Wahab M, Piazza N, Pighi M, Modolo R, van Mourik M, Wykrzykowska J, de Winter RJ, Lemos PA, de Brito FS Jr, Kawashima H, Søndergaard L, Rosseel L, Wang R, Gao C, Tao L, Rück A, Kim WK, van Royen N, Terkelsen CJ, Nissen H, Adam M, Rudolph TK, Wienemann H, Torii R, Josef Neuman F, Schoechlin S, Chen M, Elkoumy A, Elzomor H, Amat-Santos IJ, Mylotte D, Soliman O, Onuma Y. Quantitative aortography for assessment of aortic regurgitation in the era of percutaneous aortic valve replacement. *Front Cardiovasc Med*. 2023;10:1161779.
7. Tsai TY, Revaiah PC, Piazza N, Onuma Y, Serruys PW. Imaging guidance in Transcatheter Aortic Valve Implantation (preprocedural and intra-procedural imaging): Angiography. In: Serruys PW, Cribier A, Elchaninoff H, Grube E, Makkar R, Onuma Y, Webb JG, Leon MB, editors. *Transcatheter Aortic Valve Implantation: Current and Future Developments*. 2nd edition Boca Raton: CRC Press; 2024. p. 104-15.
8. Vahanian A, Beyersdorf F, Praz F, Milojevic M, Baldus S, Bauersachs J, Capodanno D, Conradi L, De Bonis M, De Paulis R, Delgado V, Freemantle N, Haugaa KH, Jeppsson A, Jüni P, Pierard L, Prendergast BD, Sádaba JR, Tribouilloy C, Wojakowski W. 2021 ESC/EACTS Guidelines for the management of valvular heart disease. *EuroIntervention*. 2022;17:e1126-96.
9. Herrmann HC, Mehran R, Blackman DJ, Bailey S, Möllmann H, Abdel-Wahab M, Ben Ali W, Mahoney PD, Ruge H, Wood DA, Bleiziffer S, Ramlawi B, Gada H, Petronio AS, Resor CD, Merhi W, Garcia Del Blanco B, Attizzani GF, Batchelor WB, Gillam LD, Guerrero M, Rogers T, Rovin JD, Szerlip M, Whisenant B, Deeb GM, Grubb KJ, Padang R, Fan MT, Althouse AD, Tchétché D; SMART Trial Investigators. Self-Expanding or Balloon-Expandable TAVR in Patients with a Small Aortic Annulus. *N Engl J Med*. 2024;390:1959-71.
10. Abbas AE, Khalili H, Madanat L, Elmariah S, Shannon F, Al-Azizi K, Waggoner T, Pilgrim T, Okuno T, Bavry A, Ternacle J, Christensen J, Cabau JR, Mack M, Pibarot P. Echocardiographic Versus Invasive Aortic Valve Gradients in Different Clinical Scenarios. *J Am Soc Echocardiogr*. 2023;36:1302-14.
11. Di Pietro G, Improta R, Bruno F, De Filippo O, Leone PP, Nebiolo M, Giacobbe F, Caporusso D, Birtolo LI, Ielasi A, Mohamed AW, Ho KW, Meguro K, Ferrara J, Waksman R, Pilgrims T, McKay RG, Seiffert M, Massimo M, De Ferrari GM, D'Ascenzo F. Impact of Small Aortic Annuli

on the Performance of Transcatheter Aortic Valve Replacement Bioprostheses: An Updated Meta-Analysis of Recent Studies. *Am J Cardiol*. 2024;229:1-12.

12. Moscarelli M, Sollami G, Lentini E, Prestera R, Pernice V, Violante F, Cuffari F, Pasquale CD, La Grutta L, Grassedonio E, Speziale G, Fattouch K. Self-Expandable Prosthesis Valve Adaptation: Non-Uniform Expansion and Stent Frame Decoupling. *Am J Cardiol*. 2023;207:93-9.
13. Di Martino LFM, Soliman OII, van Gils L, Vletter WB, Van Mieghem NM, Ren B, Galema TW, Schultz C, de Jaegere PPT, Di Biase M, Geleijnse ML. Relation between calcium burden, echocardiographic stent frame eccentricity and paravalvular leakage after corevalve transcatheter aortic valve implantation. *Eur Heart J Cardiovasc Imaging*. 2017;18:648-53.
14. Lanz J, Kim WK, Walther T, Burgdorf C, Möllmann H, Linke A, Redwood S, Thilo C, Hilker M, Joner M, Thiele H, Conzelmann L, Conradi L, Kerber S, Schymik G, Prendergast B, Husser O, Stortecky S, Heg D, Jüni P, Windecker S, Pilgrim T; SCOPE I investigators. Safety and efficacy of a self-expanding versus a balloon-expandable bioprosthesis for transcatheter aortic valve replacement in patients with symptomatic severe aortic stenosis: a randomised non-inferiority trial. *Lancet*. 2019;394:1619-28.
15. Thiele H, Kurz T, Feistritz HJ, Stachel G, Hartung P, Eitel I, Marquetand C, Nef H, Doerr O, Lauten A, Landmesser U, Abdel-Wahab M, Sandri M, Holzhey D, Borger M, Ince H, Öner A, Meyer-Saracai R, Wienbergen H, Fach A, Frey N, König IR, Vonthein R, Rückert Y, Funkat AK, de Waha-Thiele S, Desch S. Comparison of newer generation self-expandable vs. balloon-expandable valves in transcatheter aortic valve implantation: the randomized SOLVE-TAVI trial. *Eur Heart J*. 2020;41:1890-9.
16. Nombela-Franco L, Rodés-Cabau J, DeLarochellière R, Larose E, Doyle D, Villeneuve J, Bergeron S, Bernier M, Amat-Santos IJ, Mok M, Urena M, Rheault M, Dumesnil J, Côté M, Pibarot P, Dumont E. Predictive factors, efficacy, and safety of balloon post-dilation after transcatheter aortic valve implantation with a balloon-expandable valve. *JACC Cardiovasc Interv*. 2012;5:499-512.
17. Kim WK, Renker M, Nef H, Möllmann H, Walther T, Choi YH, Hamm CW. Efficacy and Safety of Postdilation for a Self-Expanding Transcatheter Heart Valve. *J Invasive Cardiol*. 2022;34:E448-54.
18. Mack MJ, Leon MB, Thourani VH, Pibarot P, Hahn RT, Genereux P, Kodali SK, Kapadia SR, Cohen DJ, Pocock SJ, Lu M, White R, Szerlip M, Ternacle J, Malaisrie SC, Herrmann HC, Szezo WY, Russo MJ, Babaliaros V, Smith CR, Blanke P, Webb JG, Makkar R; PARTNER 3 Investigators. Transcatheter Aortic-Valve Replacement in Low-Risk Patients at Five Years. *N Engl J Med*. 2023;389:1949-60.
19. Popma JJ, Deeb GM, Yakubov SJ, Mumtaz M, Gada H, O'Hair D, Bajwa T, Heiser JC, Merhi W, Kleiman NS, Askew J, Sorajja P, Rovin J, Chetcuti SJ, Adams DH, Teirstein PS, Zorn GL 3rd, Forrest JK, Tchétché D, Resar J, Walton A, Piazza N, Ramlawi B, Robinson N, Petrossian G, Gleason TG, Oh JK, Boulware MJ, Qiao H, Mugglin AS, Reardon MJ; Evolut Low Risk Trial Investigators. Transcatheter Aortic-Valve Replacement with a Self-Expanding Valve in Low-Risk Patients. *N Engl J Med*. 2019;380:1706-15.
20. Willson AB, Webb JG, Labounty TM, Achenbach S, Moss R, Wheeler M, Thompson C, Min JK, Gurvitch R, Norgaard BL, Hague CJ, Toggweiler S, Binder R, Freeman M, Poulter R, Poulsen S, Wood DA, Leipsic J. 3-dimensional aortic annular assessment by multidetector computed tomography predicts moderate or severe paravalvular regurgitation after transcatheter aortic valve replacement: a multicenter retrospective analysis. *J Am Coll Cardiol*. 2012;59:1287-94.
21. Pagnesi M, Kim WK, Baggio S, Scotti A, Barbanti M, De Marco F, Adamo M, Eitan A, Estévez-Loureiro R, Conradi L, Toggweiler S, Mylotte D, Veulemans V, Søndergaard L, Wolf A, Giannini F, Maffeo D, Pilgrim T, Montorfano M, Zweiker D, Ferlini M, Kornowski R, Hildick-Smith D, Taramasso M, Abizaid A, Schofer J, Sinning JM, Van Mieghem NM, Wöhrle J, Khogali S, Van der Heyden JAS, Wood DA, Ielasi A, MacCarthy P, Brugaletta S, Hamm CW, Costa G, Testa L, Massucci M, Alarcón R, Schäfer U, Brunner S, Reimers B, Lunardi M, Zeus T, Vanhaverbeke M, Naber CK, Di Ienno L, Buono A, Windecker S, Schmidt A, Lanzillo G, Vaknin-Assa H, Arunothayaraj S, Saccocci M, Siqueira D, Brinkmann C, Sedaghat A, Ziviello F, Seeger J, Rottbauer W, Brouwer J, Buysschaert I, Jelisejevas J, Bharucha A, Regueiro A, Metra M, Colombo A, Latib A, Mangieri A. Incidence, Predictors, and Prognostic

Impact of New Permanent Pacemaker Implantation After TAVR With Self-Expanding Valves. *JACC Cardiovasc Interv*. 2023;16:2004-17.

22. Ravaux JM, Di Mauro M, Vernooy K, Van't Hof AW, Veenstra L, Kats S, Maessen JG, Lorusso R. One-year pacing dependency after pacemaker implantation in patients undergoing transcatheter aortic valve implantation: Systematic review and meta-analysis. *JTCVS Open*. 2021;6:41-55. e15.
23. Mendiz OA, Noč M, Fava CM, Gutiérrez Jaikel LA, Szejfman M, Plesković A, Gamboa P, Valdivieso LR, Gada H, Tang GHL. Impact of Cusp-Overlap View for TAVR with Self-Expandable Valves on 30-Day Conduction Disturbances. *J Interv Cardiol*. 2021;2021:9991528.
24. Sammour Y, Banerjee K, Kumar A, Lak H, Chawla S, Incognito C, Patel J, Kaur M, Abdelfattah O, Svensson LG, Tuzcu EM, Reed GW, Puri R, Yun J, Krishnaswamy A, Kapadia S. Systematic Approach to High Implantation of SAPIEN-3 Valve Achieves a Lower Rate of Conduction Abnormalities Including Pacemaker Implantation. *Circ Cardiovasc Interv*. 2021;14:e009407.
25. Revaiah PC, Tsai TY, Tobe A, Onuma Y, Chandra P, Serruys PW. Novel Transcatheter Aortic Valve Implantation systems from India. In: Serruys PW, Cribier A, Eltchaninoff H, Grube E, Makkar R, Onuma Y, Webb JG, Leon MB, editors. Transcatheter Aortic Valve Implantation: Current and Future Developments. 2nd edition Boca Raton: CRC Press; 2024. p. 334-41.
26. Jilaihawi H, Zhao Z, Du R, Staniloae C, Saric M, Neuburger PJ, Querijero M, Vainrib A, Hisamoto K, Ibrahim H, Collins T, Clark E, Pushkar I, Bamira D, Benenstein R, Tariq A, Williams M. Minimizing Permanent Pacemaker Following Repositionable Self-Expanding Transcatheter Aortic Valve Replacement. *JACC Cardiovasc Interv*. 2019;12:1796-807.
27. Mori S, Tretter JT, Toba T, Izawa Y, Tahara N, Nishii T, Shimoyama S, Tanaka H, Shinke T, Hirata KI, Spicer DE, Saremi F, Anderson RH. Relationship between the membranous septum and the virtual basal ring of the aortic root in candidates for transcatheter implantation of the aortic valve. *Clin Anat*. 2018;31:525-34.

## Supplementary data

**Supplementary Table 1.** Technical and device success of all three cohorts in the LANDMARK trial.

**Supplementary Table 2.** Reason for new PPI in Myval versus SAPIEN THV series and Myval versus Evolut THV series.

**Supplementary Table 3.** Results from the six-minute walk test of all three cohorts.

**Supplementary Table 4.** SF-12 scores of all three cohorts.

**Supplementary Table 5.** Thirty-day clinical outcomes in patients with a small aortic annulus.

**Supplementary Table 6.** Sensitivity analysis of the moderate-severe PVR rates within the Evolut THV series.

**Supplementary Table 7.** Comparison of EOA and mean pressure gradient between Myval and SAPIEN THV series based on fitting index.

**Supplementary Figure 1.** NYHA Classification over time in the three arms.

**Supplementary Figure 2.** Preprocedural annulus area as assessed by computed tomography (CT) and postprocedural effective orifice area (EOA) as assessed by echocardiography in patients with a small annulus ( $\leq 430 \text{ mm}^2$ ).

**Supplementary Figure 3.** Fitting index and PPI in the Myval versus SAPIEN THV series arms.

The supplementary data are published online at:  
<https://eurointervention.pcronline.com/doi/10.4244/EIJ-D-24-00951>



## Supplementary data

**Supplementary Table 1. Technical and device success of all three cohorts in the LANDMARK trial.**

	Population	Myval THV series n=379	Sapien THV series n=189	Evolut THV series n=188	Risk difference % Myval vs. Sapien	<i>P</i> value Myval vs. Sapien	Risk difference % Myval vs. Evolut	<i>P</i> value Myval vs. Evolut
<b>Technical success at exit from procedure room</b>	Intention-to-treat*	365/379 (96.3)	187/189 (98.9)	178/188 (94.7)	-2.6 (-5.4, 0.2)	0.13	1.6 (-2.5, 5.7)	0.50
	As treated	359/368 (97.6)	190/195 (97.4)	181/192 (94.3)	0.2 (-2.7, 3.0)	1.00	3.3 ( -0.8, 7.3)	0.08
	Per protocol	355/363 (97.8)	185/187 (98.9)	176/185 (95.1)	-1.1 (-3.7,1.4)	0.51	2.7 ( -1.2, 6.5)	0.15
<b>Device success rate at 30-day follow-up</b>	Intention-to-treat*	345/379 (91.0)	175/189 (92.6)	163/188 (86.7)	-1.6 (-6.7, 3.5)	0.64	4.3 (-1.7,10.4)	0.15
	As treated	339/368 (92.1)	178/195 (91.3)	166/192 (86.5)	0.8 (-4.4, 6.1)	0.85	5.6 (-0.3,11.6)	0.05
	Per protocol	335/363 (92.3)	173/187 (92.5)	161/185 (87.0)	-0.2 (-5.1, 4.7)	1.00	5.3 (-0.7, 11.2)	0.07

Intention-to-treat population: All patients randomised in the study. In this table, endpoints were assessed in the patients who underwent TAVI.

Per-protocol population: All patients who were randomised in the study, met all inclusion/exclusion criteria, were implanted with the assigned device, and completed the required follow-up. As-treated population: All patients who have signed informed consent, been enrolled, randomised, assigned a study device, and treated with one of the study devices (For example, if the patient is assigned to receive device A but instead receives device B, that subject will be considered in device B cohort). Patients who entered index procedure but received a device other than the study devices or is not treated by TAVI are not included in the as-treated analysis.

\*Endpoints were assessed in the patients who underwent TAVI

THV=Transcatheter heart valve; TAVI= Transcatheter aortic valve implantation.

**Supplementary Table 2. Reason for new PPI in Myval versus SAPIEN THV series and Myval versus Evolut THV series.**

<b>Reason for New PPI</b>	<b>Myval THV series n=57</b>	<b>Sapien THV series n=33</b>	<b>Evolut THV series n=32</b>	<b>P-value Myval vs Sapien</b>	<b>P-value Myval vs Evolut</b>
Incidence of PPI (%)	15.0	17.3	16.8	0.55	0.64
Complete heart block, n (%)	45 (79.0)	17 (51.5)	28 (87.5)	0.01	0.47
Left bundle branch block (LBBB), n (%)	4 (7.0)	7 (21.2)	0	0.09	0.29
Second degree AV Block, n (%)	5 (8.8)	4 (12.1)	2 (6.3)	0.72	1.00
Bradycardia, n (%)	2 (3.5)	3 (9.1)	0	0.35	0.53
Infra-hisian conduction disturbances, n (%)	0	1 (3.0)	0	0.37	-
Combination of first degree AV block and LBBB, n (%)	0	0	1 (3.1)	-	0.36
Atrial fibrillation with significant pause, n (%)	1 (1.8)	0	0	1.00	1.00
Alternating left and right bundle branch block, n (%)	0	1 (3.0)	0	0.37	-
5-second pause, n (%)	0	0	1 (3.1)	-	0.36
AV: Atrioventricular; PPI: Permanent pacemaker implantation; THV: Transcatheter heart valve					

**Supplementary Table 3. Results from the six-minute walk test of all three cohorts.**

	Baseline					30 Day Follow-up					P-value (Baseline vs. 30-day)		
Six-minute Walk Test (meters)	Myval THV Series n=351	Sapien THV Series n=176	Evolut THV Series n=174	P-value Myval vs Sapien	P-value Myval vs Evolut	Myval THV Series n=323	Sapien THV Series n=162	Evolut THV Series n=162	P-value Myval vs Sapien	P-value Myval vs Evolut	Myval	Sapien	Evolut
	270.9 ± 109.5	281.8 ± 106.8	280.6 ± 112.1	0.28	0.35	313.0 ± 115.7	318.7 ± 108.1	320.7 ± 113.9	0.59	0.48	<0.0001	<0.0001	<0.0001
Data is presented as mean±SD. THV: Transcatheter heart valve.													



**Supplementary Table 4. SF-12 scores of all three cohorts.**

	Baseline					30 Day Follow-up					P-value (Baseline vs. 30-day)		
SF-12	Myval THV Series n=360	Sapien THV Series n=181	Evolut THV Series n=179	P- value Myval vs Sapien	P- value Myval vs Evolut	Myval THV Series n=344	Sapien THV Series n=175	Evolut THV Series n=175	P-value Myval vs Sapien	P-value Myval vs Evolut	Myval	Sapien	Evolut
PCS-12	33.0 ± 11.3	33.9 ± 11.0	32.3 ± 12.0	0.35	0.54	40.7 ± 12.9	42.9 ± 12.4	40.7 ± 13.1	0.07	0.98	<0.0001	<0.0001	<0.0001
MCS-12	42.2 ± 15.0	44.2 ± 14.8	42.3 ± 15.9	0.15	0.94	47.5 ± 14.7	49.7 ± 13.2	47.9 ± 15.1	0.08	0.76	<0.0001	<0.0001	<0.0001

Data is presented as mean±SD. MCS: Mental component score; PCS: Physical component score

**Supplementary Table 5. Thirty-day clinical outcomes in patients with a small aortic annulus (ITT population).**

Events, n (%)	Myval THV Series (n=123)	Sapien THV Series (n=63)	Evolut THV Series (n=55)	Risk difference (%), (95% CI) Myval vs Sapien THV Series	p-value (Myval vs Sapien THV Series)	Risk difference (%), (95% CI) Myval vs Evolut THV Series	p-value (Myval vs Evolut THV Series)
All-cause mortality	2 (2%)	1 (2%)	1 (2%)	0 (-3.8,3.89)	1.00	-0.2 (-4.6,4.2)	1.00
All stroke	3 (2%)	3 (5%)	3 (6%)	-2.4 (-9.5,4.8)	0.41	-3.1 (-10.9,4.9)	0.37
Bleeding (type-3 and type-4)	2 (2%)	0 (0)	2 (4%)	1.6 (-1.8,5.1)	0.55	-2.0 (-8.8,4.7)	0.59
Acute kidney injury (stage-2, stage-3 and stage-4)	1 (1%)	0 (0)	1 (2%)	0.8 (-1.6,3.2)	1.00	-1.0 (-5.9,3.9)	0.52
Moderate or severe prosthetic valve regurgitation	3 (2%)	0 (0)	6 (11%)	2.4 (-1.5,6.4)	0.55	-8.5 (-18.5,1.5)	0.03
Conduction system disturbances resulting in a new permanent pacemaker	15 (12%)	8 (13%)	8 (15%)	-0.5 (-11.1,10.1)	1.00	-2.3 (-14.6,9.9)	0.85
Major vascular complications	1 (1%)	2 (3%)	4 (7%)	-2.4 (-8.2,3.5)	0.27	-6.5 (-14.8,1.9)	0.03
<b>Primary composite endpoint</b>	24 (20%)	13 (21%)	18 (33%)	-1.1 (-14.5,12.2)	1.00	-13.2 (-28.8,2.3)	0.08

In the Myval THV series arm, two patients who withdrew consent before the procedure without any known events at that time were excluded from the endpoint analysis. In the SAPIEN THV series arm, one patient was excluded after randomisation by the investigator due to rapid progression of his Alzheimer's disease. In the Evolut THV series arm, one patient who withdrew consent before the procedure without any known events at that time was excluded from the endpoint analysis.

ITT: Intention-to-treat; THV: Transcatheter heart valve.

**Supplementary Table 6. Sensitivity analysis of the moderate-severe PVR rates within Evolut THV series.**

	ITT population (n=190)	As-Treated population (n=192)	Moderate- Severe PVR (n=14)	Percentage (ITT population)	Overall Percentage
Evolut R	70	71	7	10% (7/70)	7.4% (14/190)
Evolut PRO	100	106	6	6.1% (7/115)	
Evolut PRO+	10	10	1		
Evolut FX	5	5	0		
Cross-overs / Device not implanted	5	-	-		

The overall comparison of moderate/severe PVR between Myval THV series and Evolut (including Evolut R) series showed a P-value =0.06 (Myval: 3.4% vs. Evolut: 7.4%; risk difference: -4.0; 95%CI: -8.5 to 0.6).

The sensitivity analysis for moderate/severe PVR between Myval THV series and Evolut (after excluding Evolut R) series showed a P-value of 0.28 (Myval: 3.4% vs. Evolut: 6.1%; risk difference: -2.68; 95%CI: -7.98 to 2.63).



ITT: Intention-to-treat; PVR: Prosthetic valve regurgitation.

**Supplementary Table 7. Comparison of EOA and mean pressure gradient between Myval and SAPIEN THV series based on fitting index.**


	<b>Fitting Index &lt;X and &gt;Y (Inappropriate fitting index)</b>		<b>Fitting Index between X and Y (Appropriate fitting index)</b>	
<b>Group</b>	<b>Myval THV series</b>	<b>Sapien THV series</b>	<b>Myval THV series</b>	<b>Sapien THV series</b>
<b>No. of Patients, n</b>	105	90	274	99
<b>EOA, cm<sup>2</sup> (Mean (SD))</b>	2.05 (0.55)	1.73 (0.51)	1.99 (0.63)	1.86 (0.54)
<b>p-value</b>	<0.0001		0.062	
<b>Mean aortic gradient, mmHg (Mean (SD))</b>	8.05 (3.74)	10.38 (4.67)	8.07 (3.55)	9.99 (5.24)
<b>p-value</b>	0.0001		<0.0001	

EOA: Effective orifice area



**A****B**

**C**



**Supplementary Figure 1.** NYHA Classification over time in the three arms.


A. NYHA classification over time in Myval THV series.

B. NYHA classification over time in Sapien THV series.


C. NYHA classification over time in Evolut THV series.

NYHA: New York Heart Association, THV: transcatheter heart valve


Pre-procedural Annulus Area (CT) and Post-Procedural EOA (echocardiography) with indication of cases of permanent pacemaker implantation (PPI) in patients with small annulus ( $\leq 430\text{mm}^2$ )



Pre-procedural Annulus Area (CT) and Post-Procedural EOA (echocardiography) with indication of cases of Prosthetic Valve Regurgitation (PVR,  $\geq$  moderate) in patients with small annulus ( $\leq 430\text{mm}^2$ )



**Supplementary Figure 2.** Preprocedural annulus area as assessed by computed tomography (CT) and postprocedural effective orifice area (EOA) as assessed by echocardiography in patients with a small annulus ( $\leq 430\text{ mm}^2$ ).



	Fitting index <X and >Y		Fitting index between X and Y	
Group	Myval THV series	Sapien THV series	Myval THV series	Sapien THV series
No. of patients	105	90	274	99
PPI	16 (15.2)	19 (21.1)	41 (15.0)	14 (14.1)

**P=0.38**

**P=0.97**

**Supplementary Figure 3.** Fitting index and PPI in the Myval versus SAPIEN THV series arms.

PPI: Permanent pacemaker implantation